



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	F	ILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO	
10/025,977		12/21/2001	Hani Shennib	CATT-001	7816	
24353	7590	08/24/2004		EXAMINER		
	•	D & FRANCIS LL	LANDREM, KAMRIN R			
200 MIDDL SUITE 200	EFIELD I	KD		ART UNIT	PAPER NUMBER	
MENLO PARK, CA 94025				3738		
				DATE MAILED: 08/24/200	DATE MAILED: 08/24/2004	

Please find below and/or attached an Office communication concerning this application or proceeding.

	Application No.	Applicant(s)	- 17			
Advisory Action	10/025,977	SHENNIB ET AL.	₩			
, lav. 100. y / louio.	Examiner	Art Unit				
	Kamrin R. Landrem	3738				
The MAILING DATE of this communication appe	ears on the cover sheet with the c	orrespondence add	ress			
THE REPLY FILED 02 July 2004 FAILS TO PLACE TH Therefore, further action by the applicant is required to a final rejection under 37 CFR 1.113 may only be either: ('condition for allowance; (2) a timely filed Notice of Appe	void abandonment of this applic 1) a timely filed amendment whi	cation. A proper re ch places the appli	ply to a cation in			
PERIOD FOR RE	EPLY [check either a) or b)]		•			
a) The period for reply expiresmonths from the mailing of	date of the final rejection.					
 The period for reply expires on: (1) the mailing date of this Adverse, will the statutory period for reply expire later the ONLY CHECK THIS BOX WHEN THE FIRST REPLY WAS 706.07(f). 	an SIX MONTHS from the mailing date of FILED WITHIN TWO MONTHS OF THI	f the final rejection. E FINAL REJECTION. S	See MPEP			
Extensions of time may be obtained under 37 CFR 1.136(a). The data have been filed is the date for purposes of determining the period of extens 7 CFR 1.17(a) is calculated from: (1) the expiration date of the shortened b) above, if checked. Any reply received by the Office later than three most patent term adjustment. See 37 CFR 1.704(b).	sion and the corresponding amount of the d statutory period for reply originally set in	fee. The appropriate ex the final Office action; or	tension fee under (2) as set forth in			
1 A Notice of Appeal was filed on Appellant'37 CFR 1.192(a), or any extension thereof (37 CF						
The proposed amendment(s) will not be entered b	ecause:					
(a) \square they raise new issues that would require furth	er consideration and/or search (see NOTE below);				
(b) they raise the issue of new matter (see Note I						
(c) ☐ they are not deemed to place the application issues for appeal; and/or	in better form for appeal by mat	erially reducing or	simplifying the			
(d) They present additional claims without cancel	ling a corresponding number of	finally rejected clai	ms.			
NOTE:						
3. Applicant's reply has overcome the following reject	ction(s):					
 Newly proposed or amended claim(s) would canceling the non-allowable claim(s). 	l be allowable if submitted in a s	eparate, timely file	d amendment			
5. ☐ The a) ☐ affidavit, b) ☐ exhibit, or c) ☐ reconstruction in condition for allow 6. ☐ The affidavit or exhibit will NOT be considered be	vance because: See Continuation	<u>Sheet</u> .				
raised by the Examiner in the final rejection.	t(a) a) will not be entored as b	N□ will be entered	and an			
7. For purposes of Appeal, the proposed amendmen explanation of how the new or amended claims w			and an			
The status of the claim(s) is (or will be) as follows:						
Claim(s) allowed:						
Claim(s) objected to:						
Claim(s) rejected: <u>9-13,15-18 and 30-38</u> .						
Claim(s) withdrawn from consideration:						
8. \square The drawing correction filed on is a) \square app	proved or b) disapproved by	the Examiner.				
9. Note the attached Information Disclosure Stateme	ent(s)(PTO-1449) Paper No(s).	·				
10. Other:	Qaylo J. Isabella	Kamrin Landrem				
	Primary Examiner	Examiner AU 3738				

Continuation of 5. does NOT place the application in condition for allowance because: The applicant's arguments are unpersuasive. Contrary to applicant's arguments Kuehn does in fact disclose an apparatus comprising a fastener 402, which can be used to both temporarily grasp and permanently secure (via tack 424) together the leaflets of a cardiac valve. (See Figures 16-17 and column 8, line 59- column 9 line 45). The fastener applicator 404 is an instrument that is used to apply the permanent fastening means 424 and the instrument is then removed from the body. Grasping arms 402 are capable of being released and removed from grasping engagement with the leaflets.